Single use trocar site closure device
Instruction forUse

【Product Name】Single use trocar site closure device

【ModelNo.】WPFB-1.4、WPFB-1.6、WPFB-1.8、WPFB-2

【Structure】

Disposable abdominal wall stapler (hereinafter referred to as "stapling device") consists of a handle, a puncture needle, a button, a clamp, and a cap.



1. Puncture needle 2. Handle 3. Cap 4. Button

Figure 1 Schematic diagram of the structure of the stapler

【mainMaterial】

Button, handle, cap (ABS), puncture needle (06Cr19Ni10).

【Application】

It is suitable for suture grasping in laparoscopic surgery to realize abdominal wall incision or poke hole suture..

【Population】

All persons

【Steps ofuse】

1. According to the needs of surgery, select the product with the appropriate specifications, open the product package, and take out the stapler.

2. Aim the needle tip of the stapler at the site to be pierced, and then pierce slowly.

3. Hold the handle of the stapler, push the button, use the clamp to grab the desired suture (2-0 and 0 sutures are recommended), release the button so that the suture is received into the needle cannula.

4. Under the monitoring of laparoscopy, the needle is punctured from the subcutaneous edge of the abdominal wall into the abdominal cavity through a full-thickness abdominal wall, and the clamp is pushed out to release the suture into the abdominal cavity; the second time without the line is placed on the contralateral subcutaneous edge of the incision Puncture the full-thickness abdominal wall into the abdominal cavity; under laparoscopic monitoring, the suture thread left in the body for the first time is clamped with a clamp, and then pulled out of the body; the two thread ends are knotted according to the standard knotting method to close the abdominal wall. incision.

【Contraindication】

Those unable to perform anesthesia and laparoscopic surgery.

【Cautions】

1.The product has been sterilized with ethylene oxide. Please do not use it if the package isdamaged.

2.The product shall be used for one time, shall not be reused, and forbidden to be

reusedafterreserialization.

3.The products are provided to professional doctors with certain basis knowledge and doctorsafter professional technicaltraining.

4.Please read this instruction for uses carefully before using the product. Please destroy theproduct afteruse.

5.When using this product, the sterile operation specification shall be strictly implemented andthe product is only for one-time use, destruction afteruse.

6. Do not use if the package is broken or expired beforeuse.

7. Please check the wafer indicator on the sterilization package of this product. "red" means thatthis product is not sterilized by ethylene oxide and cannot be directlyused.

"Green" indicates that the product has been sterilized with ethylene oxide and is ready for clinicaluse.

8. To avoid adverse consequences, please read the instructions carefully beforeuse.

9. Please check whether the product is in the valid period before use. The sterilization period isthree years. The product beyond the valid period is strictly prohibited to be used and sterilizedagain.

【Explanation of marks, sighs andabbreviations】

 KEEPDRY	 FRAGILE, HANDLE	UPWARDS	CAUTION
 USE-BY DATE	 MANUFACTURER	 DATE OFMANUFACTU	 BATCHCODE
 DONOT RE-USE	 DO NOT USEIF PACKAGEIS DAMAGED	 STERILIZED USING ETHYLENE OXIDE	 STERILE

【Storage】

The product should be stored in a dry room with excellent ventilation and without caustic gases.The relative humidity of the room must be less than80%.

【Transportation】

During transport, the product should avoid squeezing and rain. Handel with care and do notinvert.

【Sterilization method】 Ethylene oxide sterilization is used for theproduct.

【Sterilization period of validity】 The sterilization period of validity is 3 years.

【Manufacturer】

Changzhou Weipu Medical Devices Co.,Ltd

Add: NO.1-5, Chuangxin Avenue, Xinbei District, Changzhou,Jiangsu,China Tel:0519-69882737

【AuthorizedRepresentative】 EC REP

Shanghai International Holding Corp. GmbH(Europe) Add:

Eiffestrasse 80, 20537Hamburg，Germany

Tel:+49-40-2513175 Fax: +49-40-255726

Dimdi No. :DE/0000040627 E-mail:shholding@hotmail.com